



Human UCHL1 ELISA Kit

User's Manual

For Research Use Only, Not for use in diagnostic procedures

ELISA Kit for Measuring Human UCHL1

CircuLex Human UCHL1 ELISA Kit

Cat# CY-8092

Intended Use.....	1
Storage.....	1
Introduction.....	2
Principle of the Assay.....	2-3
Materials Provided.....	3
Materials Required but not Provided.....	4
Precautions and Recommendations.....	5
Sample Collection and Storage.....	6
Detailed Protocol.....	7-8
Calculations.....	9
Measurement Range.....	9
Troubleshooting.....	9
Reagent Stability.....	9
Assay Characteristics.....	10-12
Example of Test Results.....	13
References.....	14
Related Products.....	14

Intended Use

The CycLex Research Product **CircuLex Human UCHL1 ELISA Kit** is used for the quantitative measurement of human UCHL1 in cell lysate, cell culture supernatant and other biological media.

This assay kit is for research use only and not for use in diagnostic or therapeutic procedures.

Storage

- Upon receipt store all components at 4°C.
- Don't expose reagents to excessive light.

Introduction

The UCHL1, is also called PARK5 or neuronal-specific protein gene product 9.5, is a carboxyl-terminal ubiquitin hydrolase regulating ubiquitin dependent signaling pathways, recently suggested as a tumor suppressor. UCHL1 is expressed predominantly in neurons (1), representing 1 to 2% of total soluble brain protein (2), as well as in testis and ovary. In vivo, UCHL1 has been shown to be involved in the regulation of the ubiquitin pool, apoptosis, learning and memory, and its absence in mice because of spontaneous intragenic deletions yields phenotypes with neurological defects (3).

Mutations in UCHL1 have been discovered in a German family with Parkinson disease (PD); a point mutation near the active site that changes Ile 93 to Met (I93M), which caused a partial loss of the catalytic activity of this thiol protease. This mutation has been linked to an increased risk of developing an autosomal-dominant form of PD (4). Based on the abundance in the CNS, UCHL1 has been proposed as a candidate biomarker for brain injury and ischemic strokes. It was demonstrated that UCHL1 was released from injured neurons and flow into the cerebrospinal fluid and eventually into circulating blood (5).

Principle of the Assay

The CycLex Research Product **CircuLex Human UCHL1 ELISA Kit** employs the quantitative sandwich enzyme immunoassay technique. An antibody specific for human UCHL1 is pre-coated onto a microplate. Standards and samples are pipetted into the wells and the immobilized antibody binds any human UCHL1 present. After washing away any unbound substances, an HRP conjugated antibody specific for human UCHL1 is added to the wells. Following a wash to remove any unbound antibody HRP conjugate, the remaining conjugate is allowed to react with the substrate H₂O₂-tetramethylbenzidine. The reaction is stopped by addition of acidic solution and absorbance of the resulting yellow product is measured at 450 nm. The absorbance is proportional to the concentration of human UCHL1. A standard curve is constructed by plotting absorbance values versus human UCHL1 concentrations of calibrators, and concentrations of unknown samples are determined using this standard curve.

For Research Use Only, Not for use in diagnostic procedures

Summary of Procedure

Add 100 µL of diluted samples to the wells
↓ Incubate for 1 hour at room temp.
Wash the wells
↓
Add 100 µL of HRP conjugated anti-human UCHL1 antibody
↓ Incubate for 1 hour at room temp.
Wash the wells
↓
Add 100 µL of Substrate Reagent
↓
Add 100 µL of Stop Solution
↓
Measure absorbance at 450 nm

Materials Provided

All samples and standards should be assayed in duplicate. The following components are supplied and are sufficient for the one 96-well microplate kit.

Microplate: One microplate supplied ready to use, with 96 wells (12 strips of 8-wells) in a foil, zip-lock bag with a desiccant pack. Wells are coated with anti-human UCHL1 antibody as a capture antibody.

10X Wash Buffer: One bottle containing 100 mL of 10X buffer containing 2% Tween®-20

Dilution Buffer: One bottle containing 50 mL of 1X buffer; use for reconstitution of Human UCHL1 Standard and sample dilution. Ready to use.

Human UCHL1 Standard: One vial containing 100 ng of lyophilized recombinant human UCHL1.

HRP conjugated Detection Antibody: One bottle containing 12 mL of HRP (horseradish peroxidase) conjugated anti-human UCHL1 antibody. Ready to use.

Substrate Reagent: One bottle containing 20 mL of the chromogenic substrate, tetra-methylbenzidine (TMB). Ready to use.

Stop Solution: One bottle containing 20 mL of 1 N H₂SO₄. Ready to use.

Materials Required but not Provided

- **Pipettors:** 2-20 µL, 20-200 µL and 200-1000 µL precision pipettors with disposable tips
- **Precision repeating pipettor**
- **Orbital microplate shaker**
- **Microcentrifuge and tubes** for sample preparation
- **Vortex mixer**
- **Microplate washer:** optional (Manual washing is possible but not preferable)
- **Plate reader:** capable of measuring absorbance in 96-well plates at dual wavelengths of 450/540 nm. Dual wavelengths of 450/550 or 450/595 nm can also be used. The plate can also be read at a single wavelength of 450 nm, which will give a somewhat higher reading.
- **Software package facilitating data generation and analysis :**optional
- **500 or 1000 mL graduated cylinder**
- **Reagent reservoirs**
- **Deionized water of the highest quality**
- **Disposable paper towels**

Precautions and Recommendations

- Allow all the components to come to room temperature before use.
- All microplate strips that are not immediately required should be returned to the zip-lock pouch, which must be carefully resealed to avoid moisture absorption.
- Do not use kit components beyond the indicated kit expiration date.
- Use only the microtiter wells provided with the kit.
- Rinse all detergent residues from glassware.
- Use deionized water of the highest quality.
- Do not mix reagents from different kits.
- The buffers and reagents in this kit may contain preservatives or other chemicals. Care should be taken to avoid direct contact with these reagents.
- Do not mouth pipette or ingest any of the reagents.
- Do not smoke, eat, or drink when performing the assay or in areas where samples or reagents are handled.
- Dispose of tetra-methylbenzidine (TMB) containing solutions in compliance with local regulations.
- Avoid contact with the acidic Stop Solution and Substrate Solution, which contains hydrogen peroxide.
- Wear gloves and eye protection when handling immunodiagnostic materials and samples of human origin, and these reagents. In case of contact with the Stop Solution and the Substrate Solution, wash skin thoroughly with water and seek medical attention, when necessary.
- **Biological samples may be contaminated with infectious agents. Do not ingest, expose to open wounds or breathe aerosols. Wear protective gloves and dispose of biological samples properly.**
- **CAUTION: Sulfuric Acid is a strong acid. Wear disposable gloves and eye protection when handling Stop Solution.**

Sample Collection and Storage

Cell lysates: Prepare cell lysates. Assay immediately or store the samples on ice for a few hours before assaying. Aliquots of the samples may also be stored at below -70°C for extended periods of time. Avoid repeated freeze-thaw cycles.

Cell culture supernatant: Remove any particulates by centrifugation and assay immediately or aliquot and store samples at below -70°C . Avoid repeated freeze-thaw cycles.

Other biological samples: Remove any particulates by centrifugation and assay immediately or aliquot and store samples at below -70°C . Avoid repeated freeze-thaw cycles.

Detailed Protocol

The CycLex Research Product **CircuLex Human UCHL1 ELISA Kit** is provided with removable strips of wells so the assay can be carried out on separate occasions using only the number of strips required for the particular determination. Since experimental conditions may vary, an aliquot of the human UCHL1 Standard within the kit should be included in each assay as a calibrator. Disposable pipette tips and reagent troughs should be used for all liquid transfers to avoid cross-contamination of reagents or samples.

Preparation of Working Solutions

All reagents need to be brought to room temperature prior to the assay. Assay reagents are supplied ready-to-use, with the exception of **10X Wash Buffer** and **Human UCHL1 Standard**.

1. Prepare a working solution of Wash Buffer by adding 100 mL of the **10X Wash Buffer** to 900 mL of deionized (distilled) water. Mix well. Store at 4°C for two weeks or -20°C for long-term storage.
2. Reconstitute **Human UCHL1 Standard** with **1.0 mL** of **Dilution Buffer**. The concentration of the human UCHL1 in vial should be **100 ng/mL**, which is referred as a **Master Standard** of human UCHL1.

Prepare Standard Solutions as follows:

Use the **Master Standard** to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 20 ng/mL standard (Std.1) serves as the highest standard. The **Dilution Buffer** serves as the zero standard (Blank).

	Volume of Standard	Dilution Buffer	Concentration
Std.1	120 µL of Master Standard (100 ng/mL)	480 µL	20 ng/mL
Std.2	300 µL of Std. 1 (20 ng/mL)	300 µL	10 ng/mL
Std.3	300 µL of Std. 2 (10 ng/mL)	300 µL	5 ng/mL
Std.4	300 µL of Std. 3 (5 ng/mL)	300 µL	2.5 ng/mL
Std.5	300 µL of Std. 4 (2.5 ng/mL)	300 µL	1.25 ng/mL
Std.6	300 µL of Std. 5 (1.25 ng/mL)	300 µL	0.63 ng/mL
Std.7	300 µL of Std. 6 (0.625 ng/mL)	300 µL	0.31 ng/mL
Blank	-	300 µL	0 ng/mL

Note: Do not use a Repeating pipette. Change tips for every dilution. Wet tip with Dilution Buffer before dispensing. Unused portions of Master Standard should be aliquoted and stored at below -70°C immediately. Avoid multiple freeze and thaw cycles.

Sample Preparation

- Cell lysates and other biological samples require 10- to 1,000-fold dilution.
- Cell culture supernatant samples require 1- to 100-fold dilution.

Standard Assay Procedure for Human UCHL1

1. Remove the appropriate number of microtiter wells from the foil pouch and place them into the well holder. Return any unused wells to the foil pouch, refold, seal with tape and store at 4°C.
2. Dilute samples with Dilution Buffer. (See "Sample Preparation" above.)
3. Pipette **100 µL** of **Standard Solutions (Std1-Std7, Blank)** and **diluted samples** in duplicates, into the appropriate wells.
4. Incubate the plate at room temperature (ca.25°C) for 1 hour, shaking at ca. 300 rpm on an orbital microplate shaker.
5. Wash 4-times by filling each well with Wash Buffer (350 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or microplate washer.
6. Add **100 µL** of **HRP conjugated Detection Antibody** into each well.
7. Incubate the plate at room temperature (ca.25°C) for 1 hour, shaking at ca. 300 rpm on an orbital microplate shaker.
8. Wash 4-times by filling each well with Wash Buffer (350 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or microplate washer.
9. Add **100 µL** of **Substrate Reagent**. Avoid exposing the microtiter plate to direct sunlight. Covering the plate with e.g. aluminum foil is recommended. Return Substrate Reagent to 4°C immediately after the necessary volume is removed
10. Incubate the plate at room temperature (ca.25°C) for 10-20 minutes, shaking at ca. 300 rpm on an orbital microplate shaker. The incubation time may be extended up to 30 minutes if the reaction temperature is below than 20°C.
11. Add **100 µL** of **Stop Solution** to each well in the same order as the previously added Substrate Reagent.
12. Measure absorbance in each well using a spectrophotometric microplate reader at dual wavelengths of 450/540 nm. Dual wavelengths of 450/550 or 450/595 nm can also be used. Read the microplate at 450 nm if only a single wavelength can be used. Wells must be read within 30 minutes of adding the Stop Solution.

Note-1: Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

Note-2: Reliable standard curves are obtained when either O.D. values do not exceed 0.25 units for the blank (zero concentration), or 3.0 units for the highest standard concentration. The plate should be monitored at 5-minute intervals for approximately 30 minutes.

Note-3: If the microplate reader is not capable of reading absorbance greater than the absorbance of the highest standard, perform a second reading at 405 nm. A new standard curve, constructed using the values measured at 405 nm, is used to determine human UCHL1 concentration of off-scale samples. The readings at 405 nm should not replace the on-scale readings at 450 nm.

Calculations

Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density. Plot the optical density for the standards versus the concentration of the standards and draw the best curve. The data can be linearized by using log/log paper and regression analysis may be applied to the log transformation. To determine the human UCHL1 concentration of each sample, first find the absorbance value on the y-axis and extend a horizontal line to the standard curve. At the point of intersection, extend a vertical line to the x-axis and read the corresponding human UCHL1 concentration. If the samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

1. The dose-response curve of this assay fits best to a sigmoidal 4-parameter logistic equation. The results of unknown samples can be calculated with any computer program having a 4-parameter logistic function. It is important to make an appropriate mathematical adjustment to accommodate for the dilution factor.
2. Most microtiter plate readers perform automatic calculations of analyte concentration. The calibration curve is constructed by plotting the absorbance (Y) of calibrators versus log of the known concentration (X) of calibrators, using the 4-parameter function. Alternatively, the logit log function can be used to linearize the calibration curve (i.e. logit of absorbance (Y) is plotted versus log of the known concentration (X) of calibrators).

Measurement Range

The measurement range is 0.31 ng/mL to 20 ng/mL. Any sample reading higher than the highest standard should be diluted with Dilution Buffer in higher dilution and re-assayed. Dilution factors need to be taken into consideration in calculating the human UCHL1 concentration.

Troubleshooting

1. The Human UCHL1 Standard should be run in duplicate, using the protocol described in the **Detailed Protocol**. Incubation times or temperatures significantly different from those specified may give erroneous results.
2. Poor duplicates, accompanied by elevated values for wells containing no sample, indicate insufficient washing. If all instructions in the **Detailed Protocol** were followed accurately, such results indicate a need for washer maintenance.
3. Overall low signal may indicate that desiccation of the plate has occurred between the final wash and addition of Substrate Reagent. Do not allow the plate to dry out. Add Substrate Reagent immediately after wash.

Reagent Stability

All of the reagents included in the CycLex Research Product **CircuLex Human UCHL1 ELISA Kit** have been tested for stability. Reagents should not be used beyond the stated expiration date. Upon receipt, kit reagents should be stored at 4°C, except the reconstituted UCHL1 Standard must be stored at below -70°C. Coated assay plates should be stored in the original foil bag sealed by the zip lock and containing a desiccant pack.


Assay Characteristics

1. Sensitivity

The limit of detection (defined as such a concentration of human UCHL1 giving absorbance higher than mean absorbance of blank* plus three standard deviations of the absorbance of blank: A blank + 3SD blank) is better than 253 pg/ml of sample.

* Dilution Buffer is pipetted into blank wells.

Typical Standard Curve



2. Precision

Intra-assay Precision (Precision within an assay)

Three samples* of known concentration were tested seven times on one plate to assess intra-assay precision.

- Intra-assay (Within-Run, n=7) CV=4.5-6.0 %

Human UCHL1 concentration (ng/mg total protein)

No.	Sample 1	Sample 2	Sample 3
1	32.7	59.9	10.3
2	31.1	54.8	9.4
3	28.8	54.6	9.0
4	30.1	61.0	10.5
5	32.6	60.0	10.4
6	31.5	58.8	9.3
7	32.5	59.9	10.0
max.	32.7	61.0	10.5
min.	28.8	54.6	9.0
mean	31.3	58.4	9.8
SD	1.5	2.6	0.6
CV(%)	4.7	4.5	6.0

*Sample: cell lysate

Inter-assay Precision (Precision between assays)

Three samples* of known concentration were tested in four separate assays to assess inter-assay precision.

- Inter-assay (Run-to-Run, n=4) CV=2.0-6.9 %


Human UCHL1 concentration (ng/mg total protein)

Assay	Sample 1	Sample 2	Sample 3
1	30.5	62.5	9.0
2	30.4	56.0	10.2
3	30.7	57.6	9.4
4	29.3	53.1	8.9
max.	30.7	62.5	10.2
min.	29.3	53.1	8.9
mean	30.2	57.3	9.4
SD	0.6	4.0	0.6
CV(%)	2.0	6.9	6.0

*Sample: cell lysate

3. Linearity


Two biological samples* were diluted with Dilution Buffer and assayed after dilution. The neat sample is set to 1 (4/4). Please note all samples including the neat sample are 10-fold diluted as stated in the Assay Procedure. The results are summarized in the figure below.



*Sample: cell lysate

Example of Test Results

Fig.1 Human UCHL1 concentrations in several cell culture supernatants



References

1. Osaka H, Wang YL, et al. (2003) Ubiquitin carboxy-terminal hydrolase L1 binds to and stabilizes monoubiquitin in neuron. Hum Mol Genet 12: 1945-58.
2. Wilkinson, K. D., Lee, K. M., Deshpande, S., Duerksen-Hughes, P., Boss, J. M., Pohl, J. (1989) The neuron-specific protein PGP 9.5 is a ubiquitin carboxyl-terminal hydrolase. Science 246: 670-672,
3. Saigoh K, Wang YL, et al. (1999) Intragenic deletion in the gene encoding ubiquitin carboxy-terminal hydrolase in gad mice. Nat Genet 23: 47-51.
4. Leroy, E., Boyer, R. et al. (1998) The ubiquitin pathway in Parkinson's disease. Nature 395: 451-452,.
5. Papa L, Akinyi L et al. (2010) Ubiquitin C-terminal hydrolase is a novel biomarker in humans for severe traumatic brain injury. Crit Care Med. 38: 138-44.

Related Products

- * CircuLex Human DJ-1/PARK7 ELISA Kit: Cat# CY-9050
- * CircuLex Human 14-3-3 Gamma ELISA Kit: Cat# CY-8082
- * CycLex Poly-Ubiquitinated Protein ELISA Kit: Cat# CY-7053
- * CycLex Poly-Ubiquitinated Protein Enrichment & Detection Kit: Cat# CY-7001
- * CycLex Proteasone Enrichment & Activity Assay Kit: Cat# CY-7002

PRODUCED BY

CycLex Co., Ltd.
1063-103 Terasawaoka
Ina, Nagano 396-0002
Japan
Fax: +81-265-76-7618
e-mail: info@cyclex.co.jp
URL: <http://www.cyclex.co.jp>

CycLex/CircuLex products are supplied for research use only. CycLex/CircuLex products and components thereof may not be resold, modified for resale, or used to manufacture commercial products without prior written approval from CycLex Co., Ltd.. To inquire about licensing for such commercial use, please contact us via email.